

University of Nebraska at Omaha DigitalCommons@UNO

Journal Articles

Department of Biomechanics

9-10-2019

Kinematic differences between professional and lay rescuers with and without the use of real-time cpr feedback

Lyra Clark

Ben Senderling University of Nebraska at Omaha, bsenderling@unomaha.edu

Jeff R. Gould

Chris Kaufman

Nicholas Stergiou University of Nebraska at Omaha, nstergiou@unomaha.edu

Follow this and additional works at: https://digitalcommons.unomaha.edu/biomechanicsarticles

Part of the Biomechanics Commons

Please take our feedback survey at: https://unomaha.az1.qualtrics.com/jfe/form/ SV_8cchtFmpDyGfBLE

Recommended Citation

Clark, L., Senderling, B., Gould, J., Kaufman, C., & Stergiou, N. (2019, September 10). Kinematic differences between professional and lay rescuers with and without use of real-time CPR feedback. Resuscitation, 142 (1 supplement), e9-e10. https://doi.org/10.1016/j.resuscitation.2019.06.033

This Article is brought to you for free and open access by the Department of Biomechanics at DigitalCommons@UNO. It has been accepted for inclusion in Journal Articles by an authorized administrator of DigitalCommons@UNO. For more information, please contact unodigitalcommons@unomaha.edu.



Kinematic differences between professional and lay rescuers with and without the use of real-time cpr feedback

Lyra Clark¹, Ben Senderling², Jeffrey Gould¹, Chris Kaufman¹, Nick Stergiou²

1 ZOLL Medical Corporation, Chelmsford, MA, USA

2 University of Nebraska at Omaha, Omaha, NE, USA

Purpose: Guideline-compliant cardiopulmonary resuscitation (CPR) performance can be achieved with training and use of real-time feedback. Kinematic differences are reported between experts and novices in various motor tasks. The aim of this pilot study was to investigate differences in kinematics between professional and lay rescuers during CPR performed on a manikin with and without feedback.

Methods: Professional (n = 5) and lay rescuers (n = 11) performed two minutes of continuous chest compressions on a manikin for two trials. Real-time CPR feedback provided by a defibrillator was disabled in the first trial and enabled in the second. CPR pads containing an accelerometer were used to calculate individual compression characteristics. Participants were instrumented for electromyography (EMG) and inertial motion capture and a motion capture marker was placed on the top hand. Paired and independent-sample t-tests and Pearson correlations were conducted in STATA 15.1.

Results: CPR feedback increased compression depth in lay rescuers (p < 0.05) to achieve guideline compliance. Lower bilateral hip range of motion (ROM) was recorded in lay rescuers compared with professionals without feedback (p < 0.05), but hip ROM was increased in lay rescuers with feedback enabled (p < 0.05). Hip ROM was associated with compression depth on both right (r = 0.61, p < 0.01) and left sides (r = 0.65, p < 0.01) for all rescuers. Greater left shoulder flexion was measured in lay rescuers both with (p < 0.05) and without feedback (p < 0.05). Lower extremity muscle coactivation indexes (CI) indicate greater hip extensor activity in professionals with feedback on both left (1.42 ± 0.17 vs. 0.87 ± 0.12 , p < 0.05) and right sides (1.33 ± 0.16 vs. 0.99 ± 0.07 , p < 0.05).

	Foodback	Depth (cm)	Rate	L Hip ROM	R Hip ROM	L Should Max	R Should Max
	Teeuback	Depth (cm)	(cc/mm)	(degrees)	luegrees	luegrees	(degrees)
Lay rescuers	disabled	$4.47 \pm 0.26^{a,b}$	121.37 ± 8.66	11.48 ± 0.89 ^{a,b}	11.65 <u>+</u> 0.91 ^{a,b}	57.54 ± 2.80 ^a	58.89 ± 1.74
	enabled	5.36 ± 0.20^{b}	113.53 ± 2.69	14.06 ± 1.06 ^b	14.46 ± 1.30 ^b	57.77 <u>+</u> 3.67ª	58.45 <u>+</u> 3.23
Professionals	disabled	6.73 ± 0.34 ^a	118.76 ± 7.52	16.16 <u>+</u> 1.94 ^a	16.76 ± 2.10 ^a	43.40 ± 4.05ª	53.85 ± 1.80
	enabled	6.01 ± 0.13	112.78 ± 4.17	14.00 <u>+</u> 1.65	14.25 ± 1.57	43.13 ± 2.23ª	49.95 ± 4.55
Data reported as mean <u>+</u> standard error							

^asignificant difference between lay rescuers and professional rescuers with same feedback conditions (p<0.05). ^bsignificant difference between feedback enabled and disabled within groups (p<0.05).

Conclusions: Real-time CPR feedback facilitated guideline-compliant compression performance and differences between professional and lay rescuers in body position and muscle activation were identified. The importance of these kinematic differences on rescuer fatigue and chest compression-generated blood flow should be investigated.